

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 05/04/2022

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:47
INTERNAL STANDARD	VERIFIED	19:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:47
BLANK TEST	0.00	19:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:48
BLANK TEST	0.00	19:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:50
BLANK TEST	0.00	19:51

