

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 07/13/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:16
INTERNAL STANDARD	VERIFIED	06:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:17
BLANK TEST	0.00	06:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:18
BLANK TEST	0.00	06:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:19
BLANK TEST	0.00	06:20

