

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/04/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:38
INTERNAL STANDARD	VERIFIED	05:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:39
BLANK TEST	0.00	05:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:40
BLANK TEST	0.00	05:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:41
BLANK TEST	0.00	05:42

