

SUPERVISOR TEST

WFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 08/11/2022

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:51
INTERNAL STANDARD	VERIFIED	21:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:52
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:53
BLANK TEST	0.00	21:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:54
BLANK TEST	0.00	21:55

