

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/14/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:00
INTERNAL STANDARD	VERIFIED	19:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:00
BLANK TEST	0.00	19:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:01
BLANK TEST	0.00	19:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:03
BLANK TEST	0.00	19:04

