

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 10/12/2022

OPERATOR NAME:  
NOAH A DANQUER  
Permit #: 50812  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:11
INTERNAL STANDARD	VERIFIED	03:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:11
BLANK TEST	0.00	03:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:12
BLANK TEST	0.00	03:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:13
BLANK TEST	0.00	03:14

