

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/19/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:27
INTERNAL STANDARD	VERIFIED	19:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:28
BLANK TEST	0.00	19:29
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:30
BLANK TEST	0.00	19:31

