

SUPERVISOR TEST

WFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 11/09/2022

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:52
INTERNAL STANDARD	VERIFIED	03:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:53
BLANK TEST	0.00	03:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:54
BLANK TEST	0.00	03:55
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:55
BLANK TEST	0.00	03:56

