

SUPERVISOR TEST

WFD

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/23/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:34
INTERNAL STANDARD	VERIFIED	04:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:35
BLANK TEST	0.00	04:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:36
BLANK TEST	0.00	04:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:37
BLANK TEST	0.00	04:38

