

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 12/07/2022

OPERATOR NAME:  
NOAH A DANQUER  
Permit #: 50812  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:02
INTERNAL STANDARD	VERIFIED	00:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:03
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:05

WFO

