

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/14/2022

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:07
INTERNAL STANDARD	VERIFIED	06:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:07
BLANK TEST	0.00	06:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:08
BLANK TEST	0.00	06:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:10
BLANK TEST	0.00	06:10

