

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 12/21/2022

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:01
INTERNAL STANDARD	VERIFIED	00:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:03
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:05

