

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: Oncida County SD

Instrument Serial #: 100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
4/14/22	22030	34.02	34.03	34.03	MP4192 100197	.10	.10	.10	Eilers	Y
4/20/22	22030	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
4/27/22	22030	34.03	34.03	34.02	MP4192	.10	.10	.10	Eilers	N
5/2/22	22030	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
5/4/22	22030	34.03	34.03	34.03	MP4192	.10	.10	.10	Cornish	N
5/11/22	22030	filter wheel error							Eilers	N
6/27/22	22030	34.03	34.03	34.03	MP4192	.10	.10	.10	Sorrentino	Y
7/1/22	22100	34.02	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
7/6/22	22100	34.02	34.02	34.02	MP4192	.10	.10	.10	Eilers	N
7/13/22	22100	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
7/20/22	22100	34.02	34.02	34.02	MP4192	.10	.10	.10	Oster	N
7/27/22	22100	34.03	34.03	34.03	MP4192	.10	.10	.10	Sorrentino	N

* - Please enter a ✓ to denote simulator solution change taking place