

**Infrared Instrument  
Weekly Calibration Verification Test Record**

Department: OCSO - WFO

Instrument Serial #: 100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
10/22	21280	34.02	34.03	34.03	MP4192	.10	.10	.10	Cornish	N
10/27/21	21280	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
11/03/21	21280	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	Bottle N
11/11/21	21280	34.03	34.03	34.03	MP4192	.10	.10	.10	Quinn	N
11/18/21	21280	34.03	34.03	34.03	MP4192	.10	.10	.10	Grabel/Wor	N
11/24/21	21280	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
12/1/21	21390	34.03	34.03	34.03	MP4192	.10	.10	.10	Cornish	Y
12/8/21	21390	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
2/15/21	21390	34.03	34.03	34.03	MP4192	.10	.10	.10	Grabel/Wor	N
12/22/21	21390	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
12/29/21	21390	34.03	34.03	34.03	MP4192	.10	.10	.10	Cusack	N
1/1/22	21390	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	Bottle N

\* Please enter a ✓ to denote simulator solution change taking place

**Infrared Instrument  
Weekly Calibration Verification Test Record**

Department: \_\_\_\_\_

Instrument Serial #: \_\_\_\_\_

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
1-11-22	21390	34.02	34.03	34.03	MP4192	.10	.10	.10	Eilers	No
1-13-22	21390	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	No
1-19-22	21390	34.01	34.01	34.04	MP4192	.10	.10	.10	Eilers	No
1/26/22	21390	34.03	34.03	34.03	MP4192	.10	.10	.10	Przybylo	No
2/2/22	21440	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	Yes
2/9/22	21440	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	No
2/16/22	21440	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	No
2/24/22	21940	34.03	34.03	34.03	MP4192	.10	.10	.10	Osika	No
3/1/22	21440	34.02	34.03	34.03	MP4192	.10	.10	.10	Cornish	<del>No</del>
3/9/22	21440	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	No
3/16/22	21440	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	No
3/23/22	21440	34.02	34.03	34.03	MP4192	.10	.10	.10	Allen	No

\* - Please enter a √ to denote simulator solution change taking place