

100197

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: _____

Instrument Serial #: _____

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
10/5/22	22260	34.02	34.02	34.02	MP4192	.10	.10	.10	DANGER	N
10/12/22	22260	34.03	34.03	34.03	MP4192	.10	.10	.10	DANGER	N
10/19/22	22260	34.03	34.03	34.03	MP4192	.10	.10	.10	DANGER	N
10/26/22	22260	34.03	34.02	34.03	MP4192	.10	.10	.10	DANGER	N
11-1-22	22260	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
11-2-22	22260	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N
11/9/22	22260	34.03	34.03	34.03	MP4192	.10	.10	.10	OSIRA	N
11-18-22	22260	34.02	34.02	34.02	MP4192	.10	.10	.10	Eilers	N
11/23/22	22260	34.03	34.03	34.03	MP4192	.10	.10	.10	DANGER	N
12/1/22	22390	34.03	34.02	34.02	MP4192	.10	.10	.10	DANGER	Y
12/7/22	22390	34.03	34.03	34.03	MP4192	.10	.10	.10	DANGER	N
12/21/22	22390	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	N

* - Please enter a √ to denote simulator solution change taking place