

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/03/2023

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:04
INTERNAL STANDARD	VERIFIED	00:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:07
BLANK TEST	0.00	00:08

