

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/01/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:41
INTERNAL STANDARD	VERIFIED	17:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:41
BLANK TEST	0.00	17:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:44
BLANK TEST	0.00	17:44

