

SOLUTION CHANGE

WFO

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

Date: 10/01/2023

OPERATOR NAME:
 JASON L OSIKA
 Permit #: 35567
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:19
INTERNAL STANDARD	VERIFIED	20:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:20
BLANK TEST	0.00	20:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:21
BLANK TEST	0.00	20:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:22
BLANK TEST	0.00	20:23

