

SOLUTION CHANGE

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/01/2023

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:12
INTERNAL STANDARD	VERIFIED	02:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:13
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:14
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:15
BLANK TEST	0.00	02:16

