

SOLUTION CHANGE

WFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 03/29/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:47
INTERNAL STANDARD	VERIFIED	00:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:48
BLANK TEST	0.00	00:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:49
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:50
BLANK TEST	0.00	00:51

