

SUPERVISOR TEST

WFO

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

Date: 04/12/2023

OPERATOR NAME:
 JASON L OSIKA
 Permit #: 35567
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:07
INTERNAL STANDARD	VERIFIED	00:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:08
BLANK TEST	0.00	00:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:10
BLANK TEST	0.00	00:11

