

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/19/2023

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:34
INTERNAL STANDARD	VERIFIED	02:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:35
BLANK TEST	0.00	02:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:37
BLANK TEST	0.00	02:38

