

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 05/24/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:13
INTERNAL STANDARD	VERIFIED	04:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:13
BLANK TEST	0.00	04:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:15
BLANK TEST	0.00	04:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:16
BLANK TEST	0.00	04:17

