

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 07/19/2023

WFD

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:56
INTERNAL STANDARD	VERIFIED	17:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:57
BLANK TEST	0.00	17:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:59
BLANK TEST	0.00	18:00

