

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/16/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:41
INTERNAL STANDARD	VERIFIED	01:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:42
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:43
BLANK TEST	0.00	01:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:44
BLANK TEST	0.00	01:45

