

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/30/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:36
INTERNAL STANDARD	VERIFIED	00:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:37
BLANK TEST	0.00	00:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:39
BLANK TEST	0.00	00:40

