

SUPERVISOR TEST

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100197

*WFO*

Date: 09/21/2023

OPERATOR NAME:  
 CHRISTIAN T BRASHEAR  
 Permit #: 52127  
 Operator Agency/Dept: ONEIDA CO. S.O.  
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:55
INTERNAL STANDARD	VERIFIED	00:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:55
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:56
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:58

