

WFO

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 10/05/2023

OPERATOR NAME:  
CHRISTIAN T BRASHEAR  
Permit #: 52127  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:48
INTERNAL STANDARD	VERIFIED	01:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:49
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:51
BLANK TEST	0.00	01:52

