

SUPERVISOR TEST

wfo

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 10/25/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA COUNTY  
SHERIFF'S OFFICE  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:16
INTERNAL STANDARD	VERIFIED	02:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:17
BLANK TEST	0.00	02:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:19
BLANK TEST	0.00	02:20

