

SUPERVISOR TEST

WFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 01/13/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:46
INTERNAL STANDARD	VERIFIED	00:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:46
BLANK TEST	0.00	00:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:47
BLANK TEST	0.00	00:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:48
BLANK TEST	0.00	00:49

