

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 01/25/2023

OPERATOR NAME:  
JAY V SORRENTINO  
Permit #: 38962  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:07
INTERNAL STANDARD	VERIFIED	00:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:10
BLANK TEST	0.00	00:11

