

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/09/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:12
INTERNAL STANDARD	VERIFIED	21:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:13
BLANK TEST	0.00	21:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:14
BLANK TEST	0.00	21:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:15
BLANK TEST	0.00	21:16

