

SUPERVISOR TEST

~~WFO~~ WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/15/2023

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:10
INTERNAL STANDARD	VERIFIED	01:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:11
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:12
BLANK TEST	0.00	01:13

