

SUPERVISOR TEST

WFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 02/22/2023

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:38
INTERNAL STANDARD	VERIFIED	20:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:39
BLANK TEST	0.00	20:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:40
BLANK TEST	0.00	20:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:41
BLANK TEST	0.00	20:42

