

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/06/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:03
INTERNAL STANDARD	VERIFIED	21:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:03
BLANK TEST	0.00	21:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:04
BLANK TEST	0.00	21:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:05
BLANK TEST	0.00	21:06

