

SUPERVISOR TEST

wfo

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 11/21/2023

OPERATOR NAME:  
ROBERT J PRODY  
Permit #: 50912  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:33
INTERNAL STANDARD	VERIFIED	22:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:35
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:36
BLANK TEST	0.00	22:37

