

SUPERVISOR TEST

WFO

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

Date: 12/06/2023

OPERATOR NAME:
 JASON L OSIKA
 Permit #: 35567
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:17
INTERNAL STANDARD	VERIFIED	03:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:17
BLANK TEST	0.00	03:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:18
BLANK TEST	0.00	03:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:20
BLANK TEST	0.00	03:21

