

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/14/2023

WFO

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:19
INTERNAL STANDARD	VERIFIED	00:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:19
BLANK TEST	0.00	00:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:20
BLANK TEST	0.00	00:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:22
BLANK TEST	0.00	00:23

