

SUPERVISOR TEST

WFD

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/20/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:14
INTERNAL STANDARD	VERIFIED	01:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:15
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:16
BLANK TEST	0.00	01:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:17
BLANK TEST	0.00	01:18

