

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/27/2023

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:15
INTERNAL STANDARD	VERIFIED	18:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:15
BLANK TEST	0.00	18:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:16
BLANK TEST	0.00	18:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:18
BLANK TEST	0.00	18:18

WFO

