

WFO

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/02/2024

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:09
INTERNAL STANDARD	VERIFIED	00:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:10
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13

