

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/01/2024

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:41
INTERNAL STANDARD	VERIFIED	19:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:42
BLANK TEST	0.00	19:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:43
BLANK TEST	0.00	19:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:44
BLANK TEST	0.00	19:45

