

WFO

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/09/2024

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:40
INTERNAL STANDARD	VERIFIED	17:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:41
BLANK TEST	0.00	17:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:43
BLANK TEST	0.00	17:44

