

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/02/2024

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:00
INTERNAL STANDARD	VERIFIED	02:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:00
BLANK TEST	0.00	02:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:02
BLANK TEST	0.00	02:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:03
BLANK TEST	0.00	02:04

