

SOLUTION CHANGE

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/22/2024

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:53
INTERNAL STANDARD	VERIFIED	06:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:54
BLANK TEST	0.00	06:55
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:55
BLANK TEST	0.00	06:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:56
BLANK TEST	0.00	06:57

