

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/14/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:37
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:38
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:40
BLANK TEST	0.00	02:41

