

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/28/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:32
INTERNAL STANDARD	VERIFIED	01:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:32
BLANK TEST	0.00	01:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:34
BLANK TEST	0.00	01:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:35
BLANK TEST	0.00	01:36

