

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/10/2024

OPERATOR NAME:
JASON L OSIA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:48
INTERNAL STANDARD	VERIFIED	01:48
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	01:49
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	01:51
BLANK TEST	0.00	01:52

