

WFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/17/2024

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 5227
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:41
INTERNAL STANDARD	VERIFIED	22:41
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:42
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	22:42
BLANK TEST	0.00	22:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:44
BLANK TEST	0.00	22:44

