

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 10/31/2024

*WFO*

OPERATOR NAME:  
MATTHEW C CADWELL  
Permit #: 53014  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:00
INTERNAL STANDARD	VERIFIED	03:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:01
BLANK TEST	0.00	03:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:02
BLANK TEST	0.00	03:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:03
BLANK TEST	0.00	03:04

